

SUPERVISOR TEST

WARREN COUNTY  
SHERRIF'S OFFICE  
DATAMASTER dmt: 144506

Date: 07/18/2016

OPERATOR NAME:  
KENNETH W SMITH  
Permit #: 35604  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 16090

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:31
INTERNAL STANDARD	VERIFIED	03:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:31
BLANK TEST	0.00	03:32
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	03:33
BLANK TEST	0.00	03:33
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	03:34
BLANK TEST	0.00	03:35

*PC Kenneth W Smith*

*7/18/16*

