

SOLUTION CHANGE




WARREN COUNTY  
 SHERRIF'S OFFICE  
 DATAMASTER dmt: 144506

Date: 08/01/2016

OPERATOR NAME:  
 RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 16120

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:36
INTERNAL STANDARD	VERIFIED	02:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:36
BLANK TEST	0.00	02:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:37
BLANK TEST	0.00	02:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:38
BLANK TEST	0.00	02:39



*515 R.J. Bartlett WSP #16  
 08-01-16*