

SUPERVISOR TEST




WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 08/09/2016

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 16120

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:49
INTERNAL STANDARD	VERIFIED	01:49
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:49
BLANK TEST	0.00	01:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:51
BLANK TEST	0.00	01:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:52
BLANK TEST	0.00	01:53



Sus Ralph J Bartlett *WCSO #16-*
 08-09-16