

SUPERVISOR TEST




WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 09/12/2016

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 16120

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:19
INTERNAL STANDARD	VERIFIED	01:19
REFERENCE STANDARD TEMP	33.94c	
REFERENCE STANDARD	0.09	01:19
BLANK TEST	0.00	01:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:21
BLANK TEST	0.00	01:21
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:22
BLANK TEST	0.00	01:23



*SGT Ralph Bartlett WSO #162
 09-12-16*