

SUPERVISOR TEST




WARREN COUNTY  
SHERRIF'S OFFICE  
DATAMASTER dmt: 144506

Date: 09/25/2016

OPERATOR NAME:  
RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 16120

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:14
INTERNAL STANDARD	VERIFIED	19:14
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	19:15
BLANK TEST	0.00	19:16
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	19:16
BLANK TEST	0.00	19:17
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	19:17
BLANK TEST	0.00	19:18



SOB Rpt / Bartlett WRSO #162  
09-25-16