

SUPERVISOR TEST




WARREN COUNTY  
 SHERRIF'S OFFICE  
 DATAMASTER dmt: 144506

Date: 10/24/2016

OPERATOR NAME:  
 RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 16170

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:52
INTERNAL STANDARD	VERIFIED	01:52
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:52
BLANK TEST	0.00	01:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:53
BLANK TEST	0.00	01:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:54
BLANK TEST	0.00	01:55



*Sgt. Bartlett* *Case #162*  
 10-24-16