

SOLUTION CHANGE



 WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 11/03/2016

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 16170

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:23
INTERNAL STANDARD	VERIFIED	16:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:23
BLANK TEST	0.00	16:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	16:24
BLANK TEST	0.00	16:25
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	16:26
BLANK TEST	0.00	16:27



SG Bartlett wcs07162
 11/3/16