

SUPERVISOR TEST




WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 12/05/2016

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 16240

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:09
INTERNAL STANDARD	VERIFIED	01:09
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:10
BLANK TEST	0.00	01:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:11
BLANK TEST	0.00	01:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:12
BLANK TEST	0.00	01:13



SG Bartlett Process WCSO #16.2
 12-5-16