

SUPERVISOR TEST

WARREN COUNTY  
SHERRIF'S OFFICE  
DATAMASTER dmt: 144506

Date: 12/19/2016

OPERATOR NAME:  
KENNETH W SMITH  
Permit #: 35604  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 16240

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:59
INTERNAL STANDARD	VERIFIED	00:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:59
BLANK TEST	0.00	01:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:00
BLANK TEST	0.00	01:01
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:01
BLANK TEST	0.00	01:02

*P.O. K. W. Smith 301*

*12/19/16*

