

SOLUTION CHANGE



WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/02/2017

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16240

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:40
INTERNAL STANDARD	VERIFIED	00:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:41
BLANK TEST	0.00	00:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:42
BLANK TEST	0.00	00:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:44
BLANK TEST	0.00	00:45



SGT *Ralph J Bartlett* WESO #162
01-02-17