

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/23/2017

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16240

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:40
INTERNAL STANDARD	VERIFIED	20:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:41
BLANK TEST	0.00	20:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:42
BLANK TEST	0.00	20:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:43
BLANK TEST	0.00	20:44

R.O. K. Smith #301

1/23/17

