

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 03/06/2017

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16380

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:49
INTERNAL STANDARD	VERIFIED	22:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:49
BLANK TEST	0.00	22:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:50
BLANK TEST	0.00	22:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:52
BLANK TEST	0.00	22:53



SES *RJ Bartlett* WASC #162
3/6/17