

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 03/21/2017

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16380

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:43
INTERNAL STANDARD	VERIFIED	02:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:43
BLANK TEST	0.00	02:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:45
BLANK TEST	0.00	02:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:46
BLANK TEST	0.00	02:47



SOS Pub, Bartlett WOSO #102

03-21-17