

SUPERVISOR TEST


WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 03/27/2017

OPERATOR NAME:  
RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 16380

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:06
INTERNAL STANDARD	VERIFIED	01:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:06
BLANK TEST	0.00	01:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:07
BLANK TEST	0.00	01:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:09
BLANK TEST	0.00	01:10



SGS *RJ Bartlett* WISO #162  
03-27-17