

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 04/04/2017

OPERATOR NAME:  
RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 17040

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:59
INTERNAL STANDARD	VERIFIED	00:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:00
BLANK TEST	0.00	01:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:01
BLANK TEST	0.00	01:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:02
BLANK TEST	0.00	01:03



Sgt R J Bartlett WOSC #162  
04-04-17