

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 05/22/2017

OPERATOR NAME:
JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17040

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:16
INTERNAL STANDARD	VERIFIED	11:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:17
BLANK TEST	0.00	11:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:18
BLANK TEST	0.00	11:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:19
BLANK TEST	0.00	11:20

JM Lopez 729

