

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/03/2017

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17150

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:23
INTERNAL STANDARD	VERIFIED	23:23

Simulator Timed Out

Ch. M. G. #246

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/03/2017

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17150

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:33
INTERNAL STANDARD	VERIFIED	23:33

Simulator Timed Out

Ch. B. Geisler #201

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208


Date: 08/03/2017

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17150

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:44
INTERNAL STANDARD	VERIFIED	23:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	INTERFERENCE	23:45

Interference Detected



Charles B. Geisler #245

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208


Date: 08/03/2017

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17150

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:56
INTERNAL STANDARD	VERIFIED	23:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	INTERFERENCE	23:57

Blank Error



Charles B. Geisler #246