

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/11/2017

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17150

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:37
INTERNAL STANDARD	VERIFIED	23:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:38
BLANK TEST	0.00	23:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:39
BLANK TEST	0.00	23:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:40
BLANK TEST	0.00	23:41

Charles B. Geisler #241

