

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 09/01/2017

OPERATOR NAME:  
JOSHUA M LOPEZ  
Permit #: 34728  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 17150

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:34
INTERNAL STANDARD	VERIFIED	23:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:35
BLANK TEST	0.00	23:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:36
BLANK TEST	0.00	23:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:38
BLANK TEST	0.00	23:39

*Sgt. JML 209*

