

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/10/2017

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17220

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:20
INTERNAL STANDARD	VERIFIED	08:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:20
BLANK TEST	0.00	08:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	08:21
BLANK TEST	0.00	08:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:23
BLANK TEST	0.00	08:24

P.O. Kenneth W. Smith



10/10/17

