

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 11/16/2017

OPERATOR NAME:
JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17220

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:41
INTERNAL STANDARD	VERIFIED	00:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:42
BLANK TEST	0.00	00:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:43
BLANK TEST	0.00	00:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:45
BLANK TEST	0.00	00:46

Sgt Jan A 099

