

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 12/08/2017

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17310

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:24
INTERNAL STANDARD	VERIFIED	00:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:24
BLANK TEST	0.00	00:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:26
BLANK TEST	0.00	00:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:27
BLANK TEST	0.00	00:28

Charles B. Geisler #204

