

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 12/16/2017

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17310

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:26
INTERNAL STANDARD	VERIFIED	03:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:27
BLANK TEST	0.00	03:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:28
BLANK TEST	0.00	03:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:30
BLANK TEST	0.00	03:31

