

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 01/23/2017

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16240

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:16
INTERNAL STANDARD	VERIFIED	21:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:16
BLANK TEST	0.00	21:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:17
BLANK TEST	0.00	21:18
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	21:18
BLANK TEST	0.00	21:19

PO K W Smith #301

1/23/17

