

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 01/30/2017

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16380

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:22
INTERNAL STANDARD	VERIFIED	00:22
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:22
BLANK TEST	0.00	00:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:24
BLANK TEST	0.00	00:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:25
BLANK TEST	0.00	00:26



505 Ref 7 *[Signature]* WCSO #162
01-30-17