

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 03/06/2017

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16380

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 22:20 |
| INTERNAL STANDARD | VERIFIED | 22:20 |
| REFERENCE STANDARD TEMP | 33.98c | |
| REFERENCE STANDARD | 0.09 | 22:20 |
| BLANK TEST | 0.00 | 22:21 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.09 | 22:21 |
| BLANK TEST | 0.00 | 22:22 |
| REFERENCE STANDARD TEMP | 33.95c | |
| REFERENCE STANDARD | 0.09 | 22:23 |
| BLANK TEST | 0.00 | 22:23 |



SGS Sgt. Bartlett WOSO #166
3/6/17