

SOLUTION CHANGE




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 04/04/2017

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 17040

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:30
INTERNAL STANDARD	VERIFIED	00:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:30
BLANK TEST	0.00	00:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:31
BLANK TEST	0.00	00:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:32
BLANK TEST	0.00	00:33



565 *RJ Bartlett* WSO #162
 04-04-17