

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 05/22/2017

OPERATOR NAME:
JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17040

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:51
INTERNAL STANDARD	VERIFIED	09:51
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	09:51
BLANK TEST	0.00	09:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:52
BLANK TEST	0.00	09:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:54
BLANK TEST	0.00	09:55

J.M.L. 799

