

SOLUTION CHANGE




WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 07/11/2017

OPERATOR NAME:
 TYLER J MORSE
 Permit #: 41956
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 17120

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:38
INTERNAL STANDARD	VERIFIED	15:39
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	15:39
BLANK TEST	0.00	15:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:40
BLANK TEST	0.00	15:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:42
BLANK TEST	0.00	15:42



PS *TL* *H417*