

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 09/01/2017

OPERATOR NAME:  
JOSHUA M LOPEZ  
Permit #: 34728  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 17150

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:20
INTERNAL STANDARD	VERIFIED	20:20
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	20:20
BLANK TEST	0.00	20:21
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	20:21
BLANK TEST	0.00	20:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	20:22
BLANK TEST	0.00	20:23

*JML 299*

