

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 10/02/2017

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17220

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:52
INTERNAL STANDARD	VERIFIED	23:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:53
BLANK TEST	0.00	23:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:54
BLANK TEST	0.00	23:55
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	23:55
BLANK TEST	0.00	23:56

See Chl B. ID #248

