

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 10/22/2017

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17220

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:23
INTERNAL STANDARD	VERIFIED	22:23
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	22:23
BLANK TEST	0.00	22:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:24
BLANK TEST	0.00	22:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:26
BLANK TEST	0.00	22:27



Charles B. Geisler #2116