

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 11/07/2017

OPERATOR NAME:
JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17220

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:34
INTERNAL STANDARD	VERIFIED	11:34
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	11:35
BLANK TEST	0.00	11:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:36
BLANK TEST	0.00	11:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:37
BLANK TEST	0.00	11:38

SA JML 299

