

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 11/15/2017

OPERATOR NAME:
JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17220

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:13
INTERNAL STANDARD	VERIFIED	20:13
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	20:14
BLANK TEST	0.00	20:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:15
BLANK TEST	0.00	20:16
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	20:16
BLANK TEST	0.00	20:17

J. Lopez

