

SUPERVISOR TEST


WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 12/25/2017

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17310

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:50
INTERNAL STANDARD	VERIFIED	18:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:51
BLANK TEST	0.00	18:52
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	18:52
BLANK TEST	0.00	18:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:53
BLANK TEST	0.00	18:54



Charles B. Geisler #244