

SOLUTION CHANGE




WARREN COUNTY  
 SHERRIF'S OFFICE  
 DATAMASTER dmt: 144506

Date: 01/02/2017

OPERATOR NAME:  
 RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 16240

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:22
INTERNAL STANDARD	VERIFIED	01:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:23
BLANK TEST	0.00	01:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:24
BLANK TEST	0.00	01:25
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	01:25
BLANK TEST	0.00	01:26



SGS *RJ Bartlett* WCRS #162

01-02-17