

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/23/2017

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16240

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:18
INTERNAL STANDARD	VERIFIED	20:18
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	20:19
BLANK TEST	0.00	20:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:20
BLANK TEST	0.00	20:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:21
BLANK TEST	0.00	20:22

P.O. K. W. Smith # 301

1/23/17

