

SUPERVISOR TEST




WARREN COUNTY
SHERRIF'S OFFICE
DATAMASTER dmt: 144506

Date: 02/07/2017

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16380

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:56
INTERNAL STANDARD	VERIFIED	00:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:57
BLANK TEST	0.00	00:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:58
BLANK TEST	0.00	00:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:59
BLANK TEST	0.00	01:00



SG *Ralph Bartlett* WCSO #162
02-07-17