

SUPERVISOR TEST



WARREN COUNTY
SHERRIF'S OFFICE
DATAMASTER dmt: 144506

Date: 02/13/2017

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 16380

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:49
INTERNAL STANDARD	VERIFIED	00:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:50
BLANK TEST	0.00	00:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:51
BLANK TEST	0.00	00:52
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:52
BLANK TEST	0.00	00:53



SGS R.J. Bartlett WCRS #162
02-13-17