

SUPERVISOR TEST



WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 02/21/2017

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 16380

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:19
INTERNAL STANDARD	VERIFIED	02:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:19
BLANK TEST	0.00	02:20
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	02:20
BLANK TEST	0.00	02:21
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:21
BLANK TEST	0.00	02:22



SGT R. J. Bartlett weso #162

02-21-17