

SUPERVISOR TEST



WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 02/27/2017

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 16380

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:37
INTERNAL STANDARD	VERIFIED	00:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:37
BLANK TEST	0.00	00:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:39
BLANK TEST	0.00	00:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:40
BLANK TEST	0.00	00:41



SES R.J. Bartlett case #162
 2/27/17