

SUPERVISOR TEST




WARREN COUNTY  
 SHERRIF'S OFFICE  
 DATAMASTER dmt: 144506

Date: 03/15/2017

OPERATOR NAME:  
 RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 16380

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:18
INTERNAL STANDARD	VERIFIED	23:18
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.10	23:19
BLANK TEST	0.00	23:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:20
BLANK TEST	0.00	23:21
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.10	23:21
BLANK TEST	0.00	23:22



Sgt R.J. Bartlett WSSC #162  
 03-15-17