

SUPERVISOR TEST




WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 03/21/2017

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 16380

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:51
INTERNAL STANDARD	VERIFIED	01:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:51
BLANK TEST	0.00	01:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:52
BLANK TEST	0.00	01:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:54
BLANK TEST	0.00	01:54



Sgt. Ralph J. Bartlett WESC #162
 03-21-17