

SUPERVISOR TEST




WARREN COUNTY
SHERRIF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/22/2017

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17040

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:00
INTERNAL STANDARD	VERIFIED	05:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	05:00
BLANK TEST	0.00	05:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:01
BLANK TEST	0.00	05:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	05:03
BLANK TEST	0.00	05:03



See DM 73. 245