

SUPERVISOR TEST

WARREN COUNTY
SHERRIF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/22/2017

OPERATOR NAME:
JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17040

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:56
INTERNAL STANDARD	VERIFIED	10:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:56
BLANK TEST	0.00	10:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:58
BLANK TEST	0.00	10:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:59
BLANK TEST	0.00	11:00

JM Lopez

