

SOLUTION CHANGE




WARREN COUNTY
SHERRIF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/11/2017

OPERATOR NAME:
TYLER J MORSE
Permit #: 41959
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17120

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:41
INTERNAL STANDARD	VERIFIED	16:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	16:42
BLANK TEST	0.00	16:43
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	16:43
BLANK TEST	0.00	16:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:44
BLANK TEST	0.00	16:45



PSM #447