

SOLUTION CHANGE

WARREN COUNTY  
SHERRIF'S OFFICE  
DATAMASTER dmt: 144506

Date: 09/01/2017

OPERATOR NAME:  
JOSHUA M LOPEZ  
Permit #: 34728  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 17150

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:50
INTERNAL STANDARD	VERIFIED	22:50
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	22:50
BLANK TEST	0.00	22:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:52
BLANK TEST	0.00	22:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:53
BLANK TEST	0.00	22:54

*2 of 2*  
*JML*

