

SUPERVISOR TEST

WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 10/10/2017

OPERATOR NAME:
 KENNETH W SMITH
 Permit #: 35604
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 17220

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:00
INTERNAL STANDARD	VERIFIED	09:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:00
BLANK TEST	0.00	09:01
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	09:02
BLANK TEST	0.00	09:02
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	09:03
BLANK TEST	0.00	09:04

PO Kenneth W Smith

KWS # 301

10/10/17

