

SOLUTION CHANGE



WARREN COUNTY  
SHERRIF'S OFFICE  
DATAMASTER dmt: 144506

Date: 11/01/2017

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 17220

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:33
INTERNAL STANDARD	VERIFIED	15:33
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	15:34
BLANK TEST	0.00	15:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:35
BLANK TEST	0.00	15:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:36
BLANK TEST	0.00	15:37



*Charles B. Geisler*