

SUPERVISOR TEST

WARREN COUNTY
SHERRIF'S OFFICE
DATAMASTER dmt: 144506

Date: 11/16/2017

OPERATOR NAME:
JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 17220

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:10
INTERNAL STANDARD	VERIFIED	00:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:10
BLANK TEST	0.00	00:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:11
BLANK TEST	0.00	00:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:12
BLANK TEST	0.00	00:13

Sgt Jm R 294

